

CORRIGENDUM

IGIB/7-2NC/284/21-22(1101)

Dated: 10th March, 2022

Sub.: Our Tender Reference No. IGIB/7-2NC/284/21-22(1101) dated 04.03.2022 for supply of High Performance Liquid Chromatography System.

This is to inform that the following specification of the above Tender may be read as amended as per the details given below:

S.No	Existing Tech. Specification	Amended Tech. Specification
1.	Flow rate 0.02 to 15mL/min.	Flow rate 0.02 to 22.5mL/min.
2.	_____	10 gradient curves or more for easy method development.
3.	_____	Lamp optimization software to maximize signal and reduce noise in both UV and Visible range.
4.	Column Heater upto 100°C.	Column Heater upto 150°C.
5.	Software for running and processing data	Software for running and processing data with embedded oracle database.
6.	UPS, PC and Printer, Fraction Collector	UPS, PC and Printer, Software controlled Fraction Collector with 3way diverse valve, 14 parameters including X-Y movement and perk simulation.

The bidder should provide detailed specifications the items at Sr. No. 6 with their make and model in the bid.

The other Tech. Specification of the Tender remain un-changed.


SPO/Sr. Controller of Stores and Purchase

19/3/2022