

## अनुबंध | Contract



अनुबंध क्रमांक | Contract No: GEMC-511687738418676

अनुबंध तिथि | Generated Date : 27-Nov-2025

संगठन विवरण   Organisation Details	खरीदार विवरण   Buyer Details
प्ररूप   Type : Central Autonomous मंत्रालय   Ministry : Ministry of Science and Technology विभाग   Department : Department of Scientific and Industrial Research (DSIR) संगठन का नाम   Organisation Name : Council of Scientific and Industrial Research (CSIR) कार्यालय क्षेत्र   Office Zone : Institute Of Genomics & Integrative Biology	पद   Designation : JSA Store and Purchase संपर्क नंबर   Contact No. : - ईमेल आईडी   Email ID : prasant.yadav@csir.res.in जीएसटीआईएन   GSTIN : - पता   Address : CSIR-Institute of Genomics & Integrative Biology, Opp. Jubilee Hall, Mall Road, Delhi, NORTH DELHI, DELHI-110007, India

वित्तीय स्वीकृति विवरण   Financial Approval Detail	भुगतान प्राधिकरण विवरण   Paying Authority Details
आईएफडी सहमति   IFD Concurrence : No प्रशासनिक अनुमोदन का पदनाम   Designation of Administrative Approval : Project Leader वित्तीय अनुमोदन का पदनाम   Designation of Financial Approval : Project leader	भुगतान का तरीका   Role: PAO भुगतान का तरीका   Payment Mode: Offline पद   Designation : FAO ईमेल आईडी   Email ID : fao.igib@nic.in जीएसटीआईएन   GSTIN : - पता   Address: CSIR-IGIB, Opp. Jubilee Hall, Mall Road, Delhi, North Delhi, DELHI-110007, India

विक्रेता विवरण   Seller Details	
जेम विक्रेता आईडी   GeM Seller ID :	SOG4220007008517
कंपनी का नाम   Company Name :	EARTH SCIENTIFIC
संपर्क नंबर   Contact No. :	08860685620
ईमेल आईडी   Email ID :	earthscientific20@gmail.com
पता   Address :	First Floor, Vipin Garden, Nawada Housing Complex, Dwarka Mor, West Delhi, DELHI-110059, -
एमएसएमई पंजीकरण संख्या   MSME Registration number :	UDYAM-DL-10-0027863
जीएसटीआईएन   GSTIN:	07CPBPM2843H1ZY (R)
एमएसई सामाजिक श्रेणी   MSE Social Category :	General
एमएसई लिंग श्रेणी   MSE Gender :	Male

\*जिसके नाम के पक्ष में GST/TAX इनवॉइस पेश किया जाएगा | GST / Tax invoice to be raised in the name of - Buyer

वितरण निर्देश | Delivery Instructions : NA

उत्पाद विवरण   Product Details						
#	आइटम विवरण   Item Description	आइटम विवरण   Ordered Quantity	इकाई   Unit	इकाई मूल्य (INR)   Unit Price (INR)	कर विभाजन (INR)   Tax Bifurcation (INR)	मूल्य (INR में सभी शुल्क और कर सहित)   Price (Inclusive of all Duties and Taxes in INR)
1	उत्पाद का नाम   Product Name : iGene Labserve Biosafety Cabinet Freestanding ब्रांड   Brand : iGene Labserve ब्रांड प्रकार   Brand Type : Registered Brand कैटलॉग की स्थिति   Catalogue Status: OEM verified catalogue कैसे बेचा जा रहा है   Selling As : OEM verified Reseller श्रेणी का नाम और चतुर्थांश   Category Name & Quadrant : Biosafety Cabinet (Q2) मॉडल   Model: IG-CIA2322 एचएसएन कोड   HSN Code: HSN not specified by seller	1	pieces	274,990	NA	274,990
कुल ऑर्डर मूल्य   Total Order Value (in INR)						274,990

परोक्षी विवरण   Consignee Detail						
क्र.सं.   S.No	परोक्षी   Consignee	वस्तु   Item	लॉट नंबर   Lot No.	मात्रा   Quantity	दिनांक के बाद डिलीवरी शुरू करना है   Delivery Start After	वितरण पूरा कब तक करना है   Delivery To Be Completed By
1	पद   Designation : ASO Jagan ईमेल आईडी   Email ID : jagan.igib@nic.in संपर्क   Contact : 011-27662099- जीएसटीआईएन   GSTIN : -	iGene Labserve Biosafety Cabinet	-	1	27-Nov-2025	12-Dec-2025

पता | Address : CSIR-IGIB, Mathura Road, Opp. Sukhdev Vihar Bus Depot, New Delhi, South Delhi, DELHI-110025, India

Freestanding

### Product Specification for iGene Labserve Biosafety Cabinet Freestanding

विनिर्देश   Specification	उप-विनिर्देश   Sub-Spec	मूल्य   Value
Standards	Conformity to Certifications	CE
	Certification number and date	21ZAAI7080C INITIAL ISSUE DATE : 15/01/2021
	Conformity to Manufacturing Standards	ICMED 90001 & 13485
	Conformity to Safety Standards	IEC 60601- or BIS equivalent
Performance Parameters	Purpose	Biological safety cabinets are used to precisely control ventilation to sweep airborne particles produced during routine microbiological procedures away from the user while maintaining a controlled environment for the specimen
	Type of Configuration	Freestanding
	Class of biosafety cabinet	CLASS II(A2 TYPE)
	Type of air cleanliness class limits(air quality) in the working area	ISO 14644 - 1 and 2 (Class 100)
	Noise level in dB	65
	Nominal inflow velocity in feet per min with air volume @450 cube feet/min	110
	Nominal down flow velocity in feet per min	55
	Direction of air flow	Vertical
	Type of air flow pattern	70% air recirculated through filter 30% air exhausted
	Pressure Monitoring System with pressure drop	13-14 mm Hg pre fiber
	Material used for working area	Stainless steel 304 grade
	Material used for the cabinet area	Stainless steel 304 grade
	Type of Sterilization used in the cabinet	UV Germicidal tube
	Type of illumination used for working surface	UV tube lights with optical reflectors to ensure uniform intensity distribution of the UV radiation throughout the cabinet along with white light.
	Intensity of light in lux	≥ 900
	Type of filter used for air supply and exhaust	HEPA (High efficiency particulate air)
	Pre filter provided	Yes
	Filtration efficiency of HEPA filter	99.99% for 0.3 micron particles
	Filtration efficiency of ULPA filter	NA
	Maximum opening height of the sash (door) in mm (Put 0 if not applicable)	0
	Width of the working area in feet	3ft and more
	Depth of the working area in feet	2ft and more
	Height of the working area in feet	2ft and more
	Type of Display	LCD
	Front door	clear, toughened sliding door of 5mm thickness
	Type of power supply in VAC,Hz	100-240 VAC, 50-60 Hz
	Automatic airflow adjustment provided	Yes
	Microprocessor control of air velocities in real time with display provided	Yes
	UV Lamp provided	With separate switch
	Control system for the choice of timed UV exposure provided	Yes
	The exposure time of UV light to be controlled through timer (in mins)	5
	Casing	Epoxy coated CRCA frame
Castors provided	Yes	
Leveling jacks provided	Yes	
Locking facility provided	Yes	

	Corrosion proof interior design and easy clean	Yes
	Drain outlet with flexible hose	Yes
	Inside space to provide full length hanging with removable shelves for storage or smaller items	No
	Auto speed increase on the opening door for maintaining velocity	Yes
	The time range to be controlled by the control system for the choice of timed UV exposure (in mins)	NA
	Biosafety cabinet should have smooth interior and exterior	Yes
	Audio and visual alarms for HEPA filter failure	Yes
	Bio safety cabinet should have Provision for fitting taps	Taps for vacuum, water and non combustible gas
Additional Accessories	Magnehelic Gauge provided	1
	Hepafilter	NA
	Number of UV Lamp	1
	Gas burner	NA
Miscellaneous Parameters	User/Technical/Maintenance manuals to be supplied in English in hard and soft copy	Yes
	Installation, Demonstration and training to be provided at consignee end	Yes
	The Principal Manufacturer Must Have Direct Presence/approved service center In India	Yes
	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	1
	Calibration Certificates	As per NABL requirement
	Copies of all certifications and reports to be provided to buyer on demand at time of supplies	Yes

## ईपीबीजी विवरण | ePBG Detail

NA

## नियम और शर्तें | Terms and Conditions

### 1. Special terms and conditions- Version:2 effective from 06-01-2023

- 1.1
- Comprehensive warranty**  
 Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares.. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables .Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
  - Service centres**  
 Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address ,telephone numbers, e mails etc at time of making the supplies .It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled .Details of toll free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
  - Source of supply**  
 It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
  - Packing and Marking**  
 Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take in to consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed .Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date , brief description of goods including quantity ,. Packing list reference number , country of origin of goods and any other relevant details.
  - Spare Parts**  
 Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM . It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies .In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.

OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.

- Installation, Training, Manuals

Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignees's representatives for operating and maintaining the equipment and Supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations , training and manuals the same shall also be applicable.

- Electrical safety checking

Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.

- Software All software updates should be provided free of cost during warranty period.

## 2. General Terms and Conditions-

2.1 This contract is governed by the [General Terms and Conditions](#), conditions stipulated to this Product/Service as provided in the Marketplace.

2.2 This Contract between the Seller and the Buyer, is for the supply of the Goods and/ or Services, detailed in the schedule above, in accordance with the General Terms and Conditions (GTC) unless otherwise superseded by Goods / Services specific Special Terms and Conditions (STC) and/ or BID/Reverse Auction Additional Terms and Conditions (ATC), as applicable

2.3 All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.

**Note:** Sellers are required to raise invoices online as per the contract terms on GeM portal. Timely invoice submission is mandatory for compliances, smooth payment processing, and will also contribute to improving their ratings.

नोट: यह सिस्टम जनरेटेड फाइल है। कोई हस्ताक्षर की आवश्यकता नहीं है। इस दस्तावेज़ का प्रिंट आउट भुगतान/लेनदेन उद्देश्य के लिए मान्य नहीं है।

Note: This is system generated file. No signature is required. Print out of this document is not valid for payment/ transaction purpose.